

A HOSPITAL BED AVAILABILITY MONITORING SYSTEM USING THE INTERNET OF THINGS

Ralph Sherwin A. Corpuz, Jofer Usa, John Cedrick Obaña, Justine Nicole Borbe, Den Arwin Salazar, Thea Marie Valeza*

Technological University of the Philippines, Manila,
Philippines

* Corresponding Author's E-mail: ralphsherwin_corpuz@tup.edu.ph

Abstract: Hospitals are considered the backbone of healthcare systems, and their significance is particularly pronounced during emergencies. However, the rapid overpopulation and recent pandemics have caused tremendous shortages of hospital beds, raising major threats to public health. This developmental research study aims to help improve the management of hospital beds through the design, development, testing, and evaluation of an Internet of Things (IoT)-based monitoring system implemented using Agile Scrum methodology. Based on the results of the hypothesis testing, the authors found that the optical sensors installed in 4 hospital rooms, with varying distance readings, are consistently accurate ($F(3, 388) = 1.000, p=0.393$), hence the authors accepted the proposed null hypothesis. Moreover, the beneficiaries, composed of the public, hospital staff, nurses, and a doctor evaluated the system to be overall very good, very efficient, very suitable, reliable, very usable, and very satisfactory. The outcome of this research aims to contribute to the United Nations' Sustainable Development Goals (UN-SDG) on good health and well-being (SDG 3).

Keywords: Hospital bed, Internet of Things, information technologies, crowdsourcing, technology management

1. INTRODUCTION

Hospitals are considered the backbone of healthcare systems around the world. They play a crucial role in maintaining good public health by providing immediate care to sick or injured people; isolation facilities or venues for medical research; and other support functions to the government and communities [1, 2]. One of the measures of quality of hospital services is the satisfaction of patients, primarily on the availability of hospital beds. In one study by [3], they found that there is a negative association between the number of hospital beds vis-à-vis the satisfaction of patients. Moreover, in another study by [4], they found that the availability of hospital beds is a crucial predictor of patient satisfaction. Interestingly, the ultimate root cause of bed shortage in hospitals is due to overcrowding, which is a common dilemma in most hospitals [5, 6], particularly during pandemics [7].

Overcrowding in hospitals happens if there is a huge number of incoming patients; there is a delay in admission, and for worse, premature deaths [8]. Overcrowding in hospitals can be further attributed to various factors, such as poor bed management [9]; issues in supply

chain management [10]; poor coordination [11]; limited manpower [12] lack of a sophisticated system to monitor real-time data [13, 14], inefficient decision-making process [15], changing decisions of patients to stay [16], and prevalence of communicable diseases in hospitals [17], among others. In the Philippines, there are fewer public hospitals than private ones in the entire archipelago, which results in more than 50% of the population who cannot afford basic healthcare services [18]. Recently during the pandemic, many Filipino families struggled to find available hospitals even in nearby towns, cities, or provinces due to overcrowding. This resulted in higher death rates and brought significant distress to patients, hospital managers, health authorities, and local and national government officials [19].

In this current study, the authors proposed a system that can provide vital information on the availability of hospital beds, hence, essentially lessening the cases of overcrowding and effectively managing hospital beds. The intended outcome is to support the healthcare system in the country through the provision of a real-time, efficient, and accurate hospital bed management system. Specifically, the authors proposed to achieve the following objectives:

1. Design a hospital bed monitoring system with the following features:
 - a. Web application for patients and the public
 - b. Desktop application for hospital administrators
 - c. IoT-powered detection system for the availability of hospital beds using reflective optical sensors, microcontroller unit, Wi-Fi module, and system peripherals.
2. Develop the system using the agile scrum methodology.
3. Test the system in terms of functional accuracy through a hypothesis-testing approach.
4. Evaluate the system based on ISO 25010:2011 in terms of performance efficiency, functional suitability, reliability, usability, and satisfaction.

2.LITERATURE REVIEW

Fortunately, several technological innovations were developed to address the problem of hospital bed management. For instance, [19] developed a real-time, web app with a trigger feature, however, it lacks a warning system for the shortage of ICU and CCU beds, which may result in an influx of ICU and CCU patients. [20] also developed another web-based bed management system with a referral feature; however, it does not have a provision to add new beds. In another study by [21], they utilized an IoT beacon with mobile and web apps. However, it has some concerns in location tracking and data security. In another study by [22], they developed a web-based ICU bed information system with a comprehensive dashboard, however, it lacks an electronic medical records-archiving feature and patient regulation system. Lastly, [23] also co-developed an autonomous robot that detects hospital room nameplates using the YOLOV4 method which is deemed 94% accurate. However, its movements are limited in tight spaces, and it has a buzzing sound that creates unnecessary noise.

Moreover, numerous technological inventions have been patented that aim to monitor hospital beds. In the work of [24], they utilized a pressure sensing system, display monitor, and alarm system to monitor available hospital beds. While the said system is informative and functional overall, it is found to be efficient, it requires numerous computer peripherals

such as a handheld Personal Digital Assistant (PDA) and a nursing station computer, hence, it is expensive to set up. Meanwhile, [25] also patented an integrated patient bed system using various diagnostic tools and equipment such as fluoroscopy, x-rays, CT-scan, ultrasound, cardiac monitor, dialysis machine, blood analysis, etc., to monitor critically ill patients. While the system is comprehensive in data generation, it is costly and complicated to manage. Overall, these technological developments and patents are deemed efficient, real-time, and sophisticated. However, further work is needed to explore alternative and sustainable solutions to optimize the monitoring of hospital beds, through simplified and decentralized management of data through continuous innovation of feature sets and the Internet of Things (IoT) framework [26].

3.METHODOLOGY

3.1 System Design

The following Figure 1 illustrates the overall design of the proposed system. In terms of its software dimension, a patient or the public can access the website application through his/her smartphone or computer to view the availability of hospital beds in nearby locations. Meanwhile, hospital staff can access the admin module through a dedicated desktop application, pre-installed in a specific hospital. The desktop app is based on the Create, Read, Update, and Delete (CRUD) processes for database management, which can manage the availability of hospital beds in direct communication with the hardware prototype through the IoT infrastructure.




Figure 1. System block diagram

The system maintains a dedicated database that serves as a repository of all processed data for both web and desktop applications. Depending on the information requested by the user, the database collects the data from the desktop application and displays it to the web application. As illustrated in Figure 2, the database is composed of 2 tables, namely “admins” and “hospitals”. The “admins” table includes the login credentials, such as username and password, as well as the user ID and hospital ID. Meanwhile, the table “hospitals” contains the demographics of various hospitals enrolled in the system, such as hospital ID, hospital name, type, bed admission type, as well as latitude and longitude of the hospital address, contact number, and system settings. Moreover, the system has a hardware prototype, which is divided into two parts – the lower part, which includes the battery holder and nameplate slot, and the upper part, which includes the microcontroller unit (MCU) and 2 optical sensors.

The following Figure 3 shows the three-dimensional model of the hardware prototype. The enclosure was designed to be slightly elevated optimize the detection consistency.




Figure 2. Database Diagram Figure 3. Hardware Prototype 3D Model

Meanwhile, the following Figures 4 and 5 show the schematic and Printed Circuit Board (PCB) designs of the hardware prototype, respectively.




Figure 4. Schematic Diagram Figure 5. PCB Design

Meanwhile, the following Figures 6, 7, and 8 detail the operational flowcharts of the web app, desktop app, and hardware prototype, respectively. These flowcharts serve as guides for the public, hospital staff, and technical support towards the effective operationalization of the system.




Figure 6. Website App Operation




Figure 7. Desktop App Operation




Figure 8. Hardware Operation

3.2. Development Framework

The authors utilized the Agile Scrum methodology as the basis of the development process, as shown in Figure 9, which covers both the software and hardware dimensions of the system.




Figure 9. Agile Scrum Methodology

Initially, the authors contributed to the project backlog by specifying the objectives, and the knowledge, software, and hardware requirements of the project. Then, they conducted sprint planning to detail the objectives, related studies, patents, methodologies, and testing and evaluation procedures. Also in this phase, the authors agreed on the specific hardware and software technologies, frameworks, and feature set. Then, they divided the tasks among

all members of the team depending on their strengths and capabilities. They also agreed on the timelines, milestones, financial costs, procurement details, communication scheme, risks, and contingency plans. Afterward, the authors prepared the sprint backlog, which lists the specific tasks, deliverables, in-charge, and target deadlines for each objective. Then, they carefully executed the objectives in coordination with technical experts, hospital authorities, and beneficiaries. Initially, they prepared the designs of the hardware, followed by the software dimensions of the system, and conducted iterative development processes for both platforms. For every phase of the development process, the authors reviewed each output to ensure conformity with the project specifications. The hardware prototype was designed to have 2 optical sensors, in which each sensor was configured to represent one bed or a total of 2 beds in a room. The authors programmed each optical sensor to detect the availability of each bed through a white nameplate. As for the testing, the authors configured each optical sensor such that once blocked, it would mean that a specific bed in a room is not available. Specifically, the authors programmed each sensor with “no block”, “block left”, “block right”, and “block both” configurations to indicate “no bed available”, “1 bed on the left is available”, “1 bed on the right is available”, and “both 2 beds are available”, respectively. With test rooms 1, 2, 3, 4, and 5, the authors varied the distance readings to 1mm 5mm, 10mm, 15mm, and 16mm, respectively. The responses of the sensors were captured automatically using a script and were collated through CSV format for further statistical analysis.




Figure 10. Testing Contraption Flowchart

Subsequently, the authors tested the null hypothesis (H_0) of the study, which states that “there is no significant difference in the functional accuracy of the system in sample hospital rooms” with a 95% confidence level and 5% error. They utilized the one-way Analysis of Variance (ANOVA) technique based on the sample size of 5 hospital nameplates with 98 test cases in each room. The analysis was conducted using one-tailed test through a statistical application software Statistical Package for the Social Sciences (SPSS) version 25. The F-statistics was computed using the following equations where: F_s = F statistic; $MSTbg$ = Mean square per treatments between groups; $MSEwg$ = Mean square per error within groups; Λ_{xy} = an observation; Π_x = group total; Gt = grand total; o_x = number in group x; and o = total observations:

$$F_s = \frac{MSTbg}{MSEwg} \quad (1)$$

$$MSTbg = \sum_{x=1}^z \left(\frac{\Pi_x^2}{o_x} \right) - \frac{Gt^2}{o} \quad (2)$$

$$MSEwg = \sum_{x=1}^z \sum_{y=1}^z \Lambda_{xy}^2 - \sum_{x=1}^z \left(\frac{\Pi_x^2}{o_x} \right) \quad (3)$$

In terms of the sprint retrospective phase, the authors involved 50 patients, the public, and another 6-hospital staff to evaluate the system using the ISO 25010:2011 international standard. They utilized a survey questionnaire based on the criteria of performance efficiency, functional suitability, reliability, usability, and satisfaction. Each question has a 5 Likert scale, where 1 and 5 were the possible lowest and highest scores, respectively. The survey form has a provision to capture the comments and suggestions from the respondents. The authors then collated and analyzed the responses using a spreadsheet and ran through the SPSS v25. Lastly, the authors carefully compiled and studied all observations, insights, and results of prototyping, quantitative analysis, survey, and feedback from the respondents for the further improvement of the system in the future.

4. RESULTS AND DISCUSSION

Shown in Figure 11 are the main components of the hardware prototype, namely: (a) 2-CNY70 reflective optical sensors, which can detect the presence of the hospital room nameplate indicating that a particular bed is occupied or not; (b) Arduino UNO MCU R3, which includes the ATmega328P processor that controls the operations of the system; (c) ESP8266 ESP-01 Wi-Fi module, which is used to connect to the internet and sends bed availability data to the online database and updates the desktop and website applications used by the general public and admin, respectively; (d) 4 piece-1.5V batteries used to power the whole hardware prototype. These hardware and software components are inter-connected through an IoT architecture.




Figure 11. Hardware Components

Meanwhile, the following Figures 11a and 11b depict the actual external and internal pictures of the hardware prototype, respectively.



Figure 11a. External View




Figure 11b. Internal View

The casing is made of a 3D-printed material, with dimensions of 55.5 mm in height, 108.5 in length, and 68 mm in width. Beneath the casing is a slightly elevated panel with 17x53x61mm detection space, sufficient enough to sense the presence of a nameplate. The nameplate is manually inserted by the hospital staff in a designated slot to indicate the

In terms of the software dimensions, the system operates into 2 platforms – web and desktop applications. Illustrated in Figure 12 is the screenshot of the website application. The authors designed the web application using Figma for the user interface and developed it using Hyper Text Markup Language (HTML), Cascading Style Sheets (CSS), JavaScript, Structured Query Language (SQL), Hypertext Pre-processor (PHP), and Asynchronous JavaScript XML (AJAX). The app features an interactive map, through Leaflet Application Programming Interface (API) and other programming techniques and algorithms, which allow the user to view, zoom, drag, or pan the location of the nearest hospital with the greatest number of available beds. Moreover, the app shows hospitals based on current location and allows users to manually view specific locations based on latitude and longitude. The app also provides alternatives to users depending on personal preferences such as public or private hospitals, room types such as ward, semi-private, or private, as well as contact details of a specific hospital. The app also works offline using AJAX and automatic location tracking using Geolocation API, and it features a Nearest Location Finder algorithm, programmed using SQL and PHP, which expedites the querying and location tracking processes. Shown in Figures 13, 14, 15, and 16 are the screenshots of the actual implementation of the Leaflet API, AJAX, Geolocation API using JavaScript, and the Nearest Location Finder algorithm, respectively.






```
index.php > index.php > head
<script>
var circle = L.circle();
circle.color = 'blue';
circle.fillColor = 'blue';
circle.fillOpacity = 0.2;
var upos = null;
var target_element = null;

const successCallback = (position) => {
  upos = [position.coords.latitude, position.coords.longitude];
  document.cookie = 'lat=' + position.coords.latitude;
  document.cookie = 'long=' + position.coords.longitude;
  position_acquired();
  createlist();
};

const errorCallback = (error) => {
  document.cookie = 'locationset=false';
  // set the target element that shows the modal
}
```

Figure 15. Geo Location API Implementation Using JavaScript



```
LIMIT 15;
} else {
  $sql = "SELECT *
  FROM (
  SELECT *,
  * DEGREES(ACOS(LEAST(1.0, COS(RADIANS(p.latpoint))
  * COS(RADIANS(z.lat))
  * COS(RADIANS(p.longpoint - z.longitude))
  + SIN(RADIANS(p.latpoint))
  * SIN(RADIANS(z.lat)))))) AS distance
  FROM hospitals AS z
  JOIN ( /* these are the query parameters */
  SELECT $lat AS latpoint, $long AS longpoint,
  $radius AS radius, 111.045 AS distance_unit /* radius is in KM */
  ) AS p ON 1=1
  WHERE z.lat
```

Figure 16. Nearest Location Finder Algorithm using SQL and PHP

Meanwhile, the desktop application of the system supports both the hardware and the central database. The authors utilized Figma for the user interface and developed the application using the Microsoft #NET framework and Windows Presentation Foundation. The desktop is intended for hospital staff, which allows them to monitor and update the availability of hospital beds based on the data fed by the hardware prototype. It features real-time querying updates, configured using C# programming language. The results are then subsequently displayed on the web application for the information of the public. Shown in Figures 17a, 17b, and 18 are the screenshots of the desktop application's homepage, modification page, and the C# Programming language implementation, respectively.




Figure 17a. Desktop App Homepage





Figure 17b. Desktop App Modification



```
public class HospitalBedsSystem
{
    // ...
    public void GetHospitalBedsStatus()
    {
        // ...
    }
    public void UpdateHospitalBedsStatus()
    {
        // ...
    }
}
```

Figure 18. C# Programming Language Implementation

The authors then tested the system by focusing on the functional accuracy of the hardware prototype, through its reflective optical sensors installed in 5 nameplates. Shown in Figure 19 is the actual testing procedure conducted. Each nameplate was tested for approximately 15 minutes with an interval of less than a second.




Figure 19. Testing Setup with Screenshot

The authors found out that the true detection range of the hardware prototype is from 1mm to a maximum of 15mm only. Unfortunately, 16mm yields to no data detection, hence, the authors excluded it in further testing the H_0 . The following Table 1 shows the summary of testing from the sample test cases. In this scenario, only 12 out of 98 test cases per set of 1mm, 5mm, 10mm, 15mm, and 16mm detection reading.

Table 1. 12 over 98 Sample Test Case Results

No.	1mm		5mm		10mm		15mm		16mm	
	Expected	Actual	Expected	Actual	Expected	Actual	Expected	Actual	Expected	Actual
87	3	3	3	3	0	0	2	2	1	NaN
88	0	0	0	0	3	3	3	3	2	NaN
89	2	2	2	2	1	1	1	1	3	NaN
90	3	3	3	3	0	0	0	0	2	NaN
91	1	1	1	1	2	2	3	3	1	NaN
92	3	3	3	3	3	3	2	3	0	NaN
93	2	2	2	2	2	2	0	0	2	NaN
94	3	3	3	3	1	1	2	2	0	NaN
95	2	2	2	2	3	3	1	1	3	NaN
96	0	0	0	0	0	0	3	3	2	NaN
97	2	2	2	2	1	1	2	2	1	NaN
98	3	3	3	3	2	2	1	1	2	NaN

Furthermore, shown in Table 2 is the result of the inferential statistics. The authors found that there was no statistically significant difference in the detection accuracy of the 4 optical sensors determined by the one-way ANOVA ($F(3, 388) = 1.00, p=0.393$). Hence, the authors accepted the H_0 . This means that the optical sensors are consistent and at par with each other in detecting accurately if a specific hospital room is available, partially available, or not available.

Table 2. ANOVA Test Results

	<i>Sum of Squares</i>	<i>df</i>	<i>Mean Square</i>	<i>F</i>	<i>Sig.</i>
Between Groups	0.008	3	0.003	1.000	0.393
Within Groups	0.990	388	0.003		
Total	0.998	391			

After the testing, the authors engaged 62 respondents, composed of 50 from the public who have experienced being admitted to a hospital, 6 hospital admin staff, 5 nurses, and 1 doctor, to evaluate the system through an online survey questionnaire based on the ISO 25010:2011. Summarized in Table 3 is the result of the evaluation. Overall, the respondents perceived that the system is “very good” with an overall average rating $\bar{x} = 4.31$. This means that the system is acceptable for the projected beneficiaries. Most of the respondents provided positive reviews and insightful suggestions for the enhancement of the website.

Table 3. Evaluation Results

<i>Criteria</i>	<i>Mean</i>	<i>Descriptive Rating</i>
Performance Efficiency	4.37	Very Good
Functional Suitability	4.46	Very Good
Reliability	4.21	Very Good
Usability	4.34	Very Good
Satisfaction	4.18	Very Good
Overall Average ($\bar{x}$)	4.31	Very Good

5. CONCLUSIONS

The proposed system aims to contribute to a more efficient bed management and decision-making process among patients and hospital managers alike. Specifically, the website application can provide real-time and accurate information on bed availability in hospitals through automatic location tracking or a manual query. It provides an interactive map that allows users or the public to view the navigational route to the selected hospital, including contact details, making it more convenient for communication purposes. Meanwhile, the desktop app allows the hospital admin staff to monitor and update bed vacancies based on the data fed by the hardware prototype through the IoT architecture. It has a secured login system, and it has a provision for adding or removing device keys to recognize or unrecognize nameplates. Moreover, the use of the IoT infrastructure, locally available materials, and the agile scrum methodology are deemed effective in attaining the study's objectives. Test results show that the system is consistently accurate in detecting the availability of hospital beds, while the expected beneficiaries have well accepted the system as per ISO 25010:2011 international standard. Hence, the system is considered a viable solution to ease the overcrowding of hospitals and a proactive communication tool for potential in-patients or the public. The outcomes of the system are expected to contribute to the UN-SDG-3 on good health and well-being. Future research should focus on applying the system to more hospitals at the regional or national levels to ensure consistency and reliability.

ACKNOWLEDGEMENT

The authors would like to acknowledge the wisdom and support of the Electronics Engineering Technology Department, College of Industrial Technology, Technological University of the Philippines- Manila, led by then-Head, Prof. Gemma D. Belga, and the faculty members – Prof. Marc V. Ardie Ardiente, Prof. Dennis J. Tabucol, Prof. Hilda V. Robino, Prof. Karen K. Foronda, and Prof. Javi Ann O. Interno (+).

REFERENCES

- [1] Nilsen, P., Seing, I., Ericsson, C., Birken, S. A., Schildmeijer, K. Characteristics of successful changes in health care organizations: an interview study with physicians, registered nurses and assistant nurses. *BMC Health Services Research*, vol. 20, 2020, pp. 1-8. <https://doi.org/10.1186/s12913-020-4999-8>
- [2] Nass, S. J., Levit L. A., Gostin, L. O. (Eds.). *Beyond the HIPAA privacy rule: enhancing privacy, improving health through research*. The National Academies Press. Washington, DC, USA. 2009.
- [3] Xesfingi, S., Vozikis, A. Patient satisfaction with the healthcare system: Assessing the impact of socio-economic and healthcare provision factors. *BMC Health Services Research*, vol. 16, no. 1, 2016, pp. 1-7, <https://doi.org/10.1186/s12913-016-1327-4>
- [4] Asamrew, N. Endris, A., Tadesse, M. Level of Patient Satisfaction with Inpatient Services and Its Determinants: A Study of a Specialized Hospital in Ethiopia. *Journal of Environmental and Public Health*, vol. 2020, Article ID 2473469, 2020, pp.1-12. <https://doi.org/10.1155/2020/2473469>
- [5] Soyemi, T.S., Aborode, A.T. Shortage of hospital bed capacity and overcrowding in emergency tertiary healthcare centers in Nigeria. *Annals of Medicine and Surgery*, vol. 82, 2022, pp. <https://doi.org/10.1016/j.amsu.2022.104675>
- [6] Bahadori, M., Teymourzadeh, E., Ravangard, R., Raadabadi, M. Factors affecting the overcrowding in outpatient healthcare. *Journal of education and health promotion*, vol.6, no. 21, 2017. <https://doi.org/10.4103/2277-9531.204742>
- [7] Savioli, G., Ceresa, I.F., Gri, N., Bavestrello Piccini, G., Longhitano, Y., Zanza, C., Piccioni, A., Esposito, C., Ricevuti, G., Bressan, M.A. Emergency department overcrowding: Understanding the factors to find corresponding solutions. *Journal of personalized medicine*, vol. 12, no. 2, 2022, p.279. <https://doi.org/10.3390/jpm12020279>
- [8] Tampubolon, L., Pujiyanto. Bed management strategy for overcrowding at the emergency department: A systematic review. *KnE Life Sciences*, vol. 4, no. 9, 2018, pp. 50–59. <https://doi.org/10.18502/kl.v4i9.3557>
- [9] Jarvis, P.R.E. Improving emergency department patient flow. *Clinical and experimental emergency medicine*, vol. 3, no. 2, 2016, pp. 63-68. <https://doi.org/0.15441/ceem.16.127>
- [10] Stansbury, N., Marlow Taylor, R., Wueste, B. A quality improvement approach to early patient discharge. *Pediatric Quality & Safety*, vol.6, no.6, 2021, e497. <https://doi.org/10.1097/pq9.0000000000000497>
- [11] Rojas-García, A., Turner, S., Pizzo, E., Hudson, E., Thomas, J., Raine, R. Impact and experiences of delayed discharge: A mixed-studies systematic review. *Health Expectations : An International Journal of Public Participation in Health Care and Health Policy*, vol.21, no.1, 2021, pp. 41–56. <https://doi.org/10.1111/hex.12619>

- [12] Di Somma, S., Paladino, L., Vaughan, L., Lalle, I., Magrini, L. Magnanti, M. Overcrowding in emergency department: an international issue. *Internal and Emergency Medicine*, vol. 10, 2015, pp.171-175. <https://doi.org/10.1007/s11739-014-1154-8>
- [13] Salway, R.J., Valenzuela, R., Shoenberger, J.M., Mallon, W.K. Viccellio, A. Emergency department (ED) overcrowding: evidence-based answers to frequently asked questions. *Revista Médica Clínica Las Condes*, vol. 28, no. 2, 2017, pp.213-219. <https://doi.org/10.1016/j.rmcl.2017.04.008>
- [14] Schmidt, R., Geisler, S., Spreckelsen, C. Decision support for hospital bed management using adaptable individual length of stay estimations and shared resources. *BMC Med Inform Decision Making*, vol. 13, no. 3, 2013, pp. 1-19. <https://doi.org/10.1186/1472-6947-13-3>
- [15] Noonan, F.; O'Brien, J.; Broderick, E.; Richardson, I. Walsh, J. Hospital bed management practices: A review. *12th International Joint Conference on Biomedical Engineering Systems and Technologies (BIOSTEC 2019) – HEALTHINF*, SciTePress, Prague, Czech Republic, pp. 326-331. <https://doi.org/10.5220/0007387403260331>
- [16] Hoffman, G.J., Min, L., Liu, H.J., Mody, L., Readmission for healthcare-acquired infections: Does patient disposition matter? *Innovation in Aging*, vol. 33, no. 1 (Suppl), 2019, p.S822. <https://doi.org/10.1093/ofid/ofz360.2148>
- [17] Silva, M.E.C., Zarsuelo, M.A.M., Naria-Maritana, M.J.N., Zordilla, Z.D., Lam, H.Y., Mendoza, M.A.F., Buan, A.K.G., Nuestro, F.K.A., Rosa, J.A.D., Padilla, C.D. Policy analysis on determining hospital bed capacity in light of universal health care. *Acta Medica Philippina*, vol. 54, no. 6, 2020, pp. 668-676. <https://doi.org/10.47895/amp.v54i6.2596>
- [18] Flores, L.J.Y., Tonato, R.R., dela Paz, G.A., Ulep, V.G., 2021. Optimizing health facility location for universal health care: A case study from the Philippines. *PloS one*, vol. 16, no. 9, 2021, e0256821, p. 1-13.
- [19] Abedian, S., Bitaraf, E., Askari, M. Advantages of a web-based real-time bed-management system for hospital admission monitoring in Iran. *Building Continents of Knowledge in Oceans of Data: The Future of Co-Created eHealth* (eds. Ugon, A. et al.), European Federation for Medical Informatics (EFMI) and IOS Press, Iran, 2018, pp. 536-540.
- [20] Abedian S, Kazemi H, Riazi H, Bitaraf E. Cross hospital bed management system. *Stud Health Technological Information*. vol. 205, 2014, pp. 126-130. PMID: 25160159.
- [21] Ajmi, F., Ajmi, F., Othman S.B., Zgaya, H., Smith, G., Renard, J-M., Hammadi, S. Impact of the automation of inpatient bed management to reduce the emergency service waiting time. *MEDINFO 2021: One World, One Health – Global Partnership for Digital Innovation*, IOS Press, vol. 290, 2022, pp. 942-946, <https://doi.org/10.3233/shti220219>
- [22] De Morais Barroca Filho, I., Sampaio, S.C., Cruz, A.P., Ramalho, V.H.F., De Azevedo, J.A.R., Da Silveira, A.C. A web-based information system for the management of ICU beds during the Coronavirus outbreak. *2020 IEEE Symposium on Computers and Communications (ISCC)*, IEEE, July 2020, Rennes, [France, pp. 1-4. <https://doi.org/10.1109/ISCC50000.2020.9219626>
- [23] Cahyo, M.P.D., Utaminingrum, F. Autonomous robot system based on room nameplate recognition using YOLOv4 method on jetson nano 2GB. *JOIV: International Journal on Informatics Visualization*, vol. 6, no. 1, 2022, pp.117-123.
- [24] Salgo, P. & Golden, R. *System and method for patient monitoring*. US Patent: 20090099480A1, May 2008. URL:

<https://patentimages.storage.googleapis.com/97/8f/44/e74a25b6ca06cc/US20090099480A1.pdf>. (Visited on 05.02.2023).

- [25] DiMaio, J.M., Watson, J., Melendez, J., Moza, R., Tibbals, H. *Integrated Patient Bed System*. US Patent: 20140090173A1, December 2013. URL: <https://patentimages.storage.googleapis.com/8f/1c/5f/f3c971be18605f/US20140090173A1.pdf>. (Visited on 05.02.2023).
- [26] Kravets, O.J., Atlasov, D.I., Gorshkov, A.V., Sidorenko, E.V., Mutina, E.I., Aksenov, I.A., Redkin, Y.V. Designing the architecture of a distributed system for information monitoring of IOT and IIOT infrastructures traffic. *International Journal on Information Technologies and Security*, vol. 16, no. 1, 2024, pp. 49-56.

Information about the authors:

Dr. Ralph Sherwin A. Corpuz is a Full Professor III at the Technological University of the Philippines. He is a Regular Member of the National Research Council of the Philippines, a Member of the Fairness Opinion Board of the Department of Science and Technology-CALABARZON, and a Senior Consultant of the Development Academy. His research interests include technology management, project management, applied Artificial Intelligence, quality assurance, and management systems. In this study, he served as the overall adviser and ensured conformance of the project study with the scientific process.

Jofer S. Usa is a Software Engineer at Samsung Electronics Philippines Corporation. His research interests include software development and optimization, embedded systems, and operating systems. In this study, he developed the desktop application and hardware prototype.

John Cedrick Obaña is a Network Engineer and Data Protection Officer (DPO) at Revlv Solutions, Inc., Philippines. His research interests include the Internet of Things (IoT). In this study, he led the compliance to project requirements and assisted in the documentation.

Justine Nicole Borbe is a UI/UX Designer at Rymera Web Co., an Australian software company, and an Assistant Project Manager at a US-based Design Studio. Her research interests include market research and service design. In this study, she designed the web application and facilitated the data collection from the respondents.

Den Arwin Salazar is a Junior Developer at Seedbox Technologies Inc., Philippines. His research interest is computer engineering technology. In this study, he helped develop the web app.

Thea Marie Valeza is a UX/UI designer at Lexagle, a Singaporean legal tech startup. Her research interests include software development and the IOT. In this study, she spearheaded the project documentation, contributed to user research, and led the design of the desktop app's UI for the hardware prototype.

Manuscript received on 25 April 2024